



Calibration Record

TCON Serial # _____

(values recorded here can be used to restore a previous calibration)

Block 4	_____	_____	.	Block 4	_____	_____
Block 3	_____	_____	.	Block 3	_____	_____
Block 2	_____	_____	.	Block 2	_____	_____
Block 1	_____	_____	.	Block 1	_____	_____
	Correction at 0	Correction at 50		Correction at 0	Correction at 50	

1. View current calibration by turning power on while pressing the arrow keys.
2. If in error, select a calibration value to change using the menu keys.
3. Correct a calibration value with the cursor and arrow keys.
4. Save corrected calibration values by pressing and holding the menu key.
5. Repeat step 1 to confirm corrections.
6. Exit without re-saving changes by turning power off.

Note: Actual calibration requires a thermometer and uses a different procedure (refer to the owner's manual).



Calibration Record

TCON Serial # _____

(values recorded here can be used to restore a previous calibration)

Block 4	_____	_____	.	Block 4	_____	_____
Block 3	_____	_____	.	Block 3	_____	_____
Block 2	_____	_____	.	Block 2	_____	_____
Block 1	_____	_____	.	Block 1	_____	_____
	Correction at 0	Correction at 50		Correction at 0	Correction at 50	

1. View current calibration by turning power on while pressing the arrow keys.
2. If in error, select a calibration value to change using the menu keys.
3. Correct a calibration value with the cursor and arrow keys.
4. Save corrected calibration values by pressing and holding the menu key.
5. Repeat step 1 to confirm corrections.
6. Exit without re-saving changes by turning power off.

Note: Actual calibration requires a thermometer and uses a different procedure (refer to the owner's manual).